

# Declaration of conformity

## Batch

Declaration of conformity given by the supplier of product according to the standard EN ISO IC 17050-1

**WERTLAB**  
**GERMANY**

Establishment applied and using system of quality  
according to the DIN EN ISO 9001: 2015, certificate made out by TÜV Rheinland FRG

Sort of product : **1567 / AS BURETTE WITH STRAIGHT STOPCOCK , GLASS KEY**


Nominal volume :	<b>50</b>	<b>ml</b>
Division :	<b>0,1</b>	<b>ml</b>
Permitted volume error :	<b>± 0,05</b>	<b>ml</b>
Average value of volume:		<b>ml</b>
Standard deviation of measurement:	<b>±</b>	<b>ml</b>
Number of charge:	<b>00 . 00</b>	

Testing instrument: balance – KERN ABT 220-4M; id. number: 34/07;  
thermometer – enclosed-scale, (0 – 50) °C; 0,01 °C; id. number: 12/12.  
Used standards are in accordance to standards of state Metrology Institute.

We are declaring to our exclusive responsibility, that product, which relates to this declaration responds to the parameters of standard:  
EN ISO 385:2005

Date of edition: 00. 00. 0000

Confirmed by:

  
Manager of Quality

000000/0